


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AUM-2-17-1D*

S.N.: *270*

Date: *DEC 28 79*



*1.0 nsec/div.  
≈ 10 VOLTS/DIV*


- a) Output signal amplitude:  
*0 TO +15 VOLTS*
- b) Pulse width:  
*0 TO 2.0 nsec*
- c) Rise time:  
*≤ 0.10 nsec*
- d) Fall time:  
*≤ 0.10 nsec.*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+24 ± 1 VOLT*

*540 mA (MAX)*

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-2-P-D REPAIRED FEB 1981*  
S.N.: *270*  
Date: *FEB 24 1981*



- a) Output signal amplitude:  
*0 TO +15 V*
- b) Pulse width:  
*0.2 TO 2.0 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+24 ± 1.0 VOLT*  
*600 mA (MAX)*

*10 VOLTS/CM*  
*1.0 NSEC/CM*  
*400 MHz*