


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-1-P*

S.N.: *285*

Date: *JAN 30 80*



*~ 10 VOLTS/DIV  
• 1.0 nSEC/DIV*

- a) Output signal amplitude:  
*> 10 VOLTS (MAX)*
- b) Pulse width:  
*0 TO 4 nSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 100 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1.0 VOLT*

*240 mA*