


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AU-1-P

S.N.: 286

Date: JAN 30 80



$\approx 10 \text{ VOLTS/DIV}$
 1.0 nSEC/DIV

- a) Output signal amplitude:
 $> 10 \text{ V (MAX)}$
- b) Pulse width:
 $0 \text{ TO } 4 \text{ nSEC}$
- c) Rise time:
 $< 100 \text{ pSEC}$
- d) Fall time:
 $< 100 \text{ pSEC}$
- e) PRF:
 $0 \text{ TO } 1.0 \text{ MHz}$
- f) Jitter, stability:
 OK
- g) Prime power:

$+15 \pm 1 \text{ VOLT}$
 215 mA