


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-AV-P*
S.N.: *287*
Date: *FEB 4 1980*



1.0 nsec / DIV
~ 10 VOLTS / DIV

- a) Output signal amplitude:
> 15 VOLTS (MAX)
- b) Pulse width:
0 TO 50 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 2.0 nsec
- e) PRF:
0 TO 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT
300 mA (MAX)