


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5

S.N.: 291

Date: FEB 22 1980



a) Output signal amplitude:

≈ 15 VOLTS

b) Pulse width:

≤ 0.5 NSEC
(10% RISE)

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0$ VOLT.

110 mA