


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V*
S.N.: *293*
Date: *FEB 28 1980*



10 nsec/div
≈ 16 volts/div

- a) Output signal amplitude:
≥ 50 VOLTS
- b) Pulse width:
0 TO 100 nsec
- c) Rise time:
≤ 0.5 nsec
- d) Fall time:
≤ 2.0 nsec
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1 VOLT

150 mA