


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-S

S.N.: 296

Date: FEB 22 1980



a) Output signal amplitude:

$\geq 15$  VOLTS

b) Pulse width:

$\leq 0.5$  NSEC  
(10% RISE)

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0$  VOLT

110 mA