


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V*
S.N.: *297*
Date: *FEB 28 1980*



10 nsec/div
≈ 16 volts/div

- a) Output signal amplitude:
≥ 50 VOLTS
- b) Pulse Width:
0 TO 100 nsec
- c) Rise time:
≤ 0.5 nsec
- d) Fall time:
≤ 2.0 nsec
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime Power:

+15 ± 1 VOLT
150 mA.