

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-V-1-N*

S.N.: *316*

Date: *APRIL 11 80*

a) Output signal amplitude:

*- 10 VOLTS*

b) Pulse width:

*0.2 TO 4.0 nSEC*

c) Rise time:

*≤ 100 PSEC (10% RISE)*

d) Fall time:

*≤ 130 PSEC*

e) PRF:

*0 TO 100 MHz*



f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0 VOLT*

*230 mA*



*1.0 nSEC/cm*  
*10 VOLTS/cm.*