


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-P

S.N.: 320

Date: APRIL 29 80



a) Output signal amplitude:

+ 40 V

b) Pulse width:

0 TO 100 NSEC

c) Rise time:

< 500 PSEC (10% RISE)

d) Fall time:

< 2.0 NSEC

e) Jitter, stability:

OK

f) PRF:

OK (SEE FIG 2)

g) Prime power:

+ 15 ± 1.0 VOLT

200 mA.

1.0 NSEC/CM.

≈ 30 VOLTS/CM