


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-N*

S.N.: *323.*

Date: *APRIL 28 80*



*1.0 nsec/cm
≈ 30 VOLTS/cm.*

- a) Output signal amplitude:
- 40 V
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 500 PSEC (10% RISE)
- d) Fall time:
< 2 NSEC
- e) Jitter, stability:
OK
- f) PRF:
OK (SEE FIG 2)
- g) Prime power:
*+15 ± 1.0 VOLT
200 mA.*