


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AU-1-05

S.N.: 326

Date: APRIL 25 80



10 VOLTS/cm
1.0 nsec/cm

a) Output signal amplitude:

0 TO +10V

b) Pulse width:

0.2 TO 4.0 nsec

c) Rise time:

≤ 100 psec
(10% rise)

d) Fall time:

100 psec

e) PRF:

0 - 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 \pm 1.0 VOLT

275 mA