


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-P
S.N.: 329
Date: MAY 21 1980



10 nsec/cm
≈ 16 volts/cm

- a) Output signal amplitude:
0 - 50 VOLTS
- b) Pulse width:
0 - 100 nSEC.
- c) Rise time:
≤ 0.5 nSEC
- d) Fall time:
≤ 2.0 nSEC.
- e) PRF:
0 - 20 KHz
- f) Jitter, stability:
OK.
- g) Prime power:
+15 ± 1 VOLT
180 mA.