


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PAV17-AV-1-N-ECL*

S.N.: *333*

Date: *JUNE 19 80*



- a) Output signal amplitude:
0 TO -10 V
- b) Pulse width:
0 TO 4 n SEC
- c) Rise time:
≤ 0.10 nsec
- d) Fall time:
< 0.1 nsec
- e) PRF:
0 TO 1.0 m PR
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 nSEC/DIV

+15 ± 1.0 VOLT

370 mA MAX

h) DRIVE LEVEL:
MEECL 10000.
TRIGGER ON
-1.690 V PULSE