


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-M2-N

S.N.: 339

Date: JUNE 13 80



50 VOLTS/DIV
5.0 NSEC/DIV

a) Output signal amplitude:

≥ 60 VOLTS

b) Pulse width:

≈ 2 NSEC

c) Jitter, stability:

OK

d) Prime power:

$+135 \pm 5.0$ VOLT

5 mA MAX.

e) TRIGGER DELAY

≤ 2.0 NSEC.

(UNIT WILL TRIGGER
FROM $+5.0$ VOLT
600 PSEC WIDE
IMPULSE)