

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVL-AV-1-P

S.N.:

340

Date:

JUNE 13 80



100 VOLT / DIV
10 nSEC / DIV

- a) Output signal amplitude;
+10 TO +100 V
- b) Pulse width:
0 TO 100 nSEC
- c) Rise time:
≤ 1.0 nSEC
- d) Fall time:
≈ 2.0 nSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT

300 mA max