


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-17

(A) S.N.: 345

Date: JUNE 30 80



a) Output signal amplitude:  
    > 15 VOLT

b) Pulse width:  
    ≤ 1.0 nsec

c) Jitter, stability:  
    OK

d) Prime power:  
    15 ± 1.0 V  
    200 mA

≈ 16 VOLTS/cm  
1.0 nsec/cm