


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P-D*

S.N.: *349*

Date: *JUNE 30 80*



*10 VOLT/cm
1.0 nsec/cm
4.0 MHz*

- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 2.0 nsec
- c) Rise time:
< 100 psec
- d) Fall time:
< 135 psec
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*+24 ± 1.0 VOLT
600 mA (MAX)*

b) DELAY

OK
*0 TO 5.0 nsec
VARIABLE*