


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-1-05-P-TTL

S.N.: 355

Date: July 15 80



- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
0 TO 4 nsec
- c) Rise time:
 ≤ 100 psec
- d) Fall time:
 ≤ 135 psec
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

1.0 nsec/div
10 volts/div

+15 \pm 1.0 VOLT
240 mA (max)