


IMPULSE GENERATOR
PERFORMANCE CHECK

Model: *AUH-MPT-2*

S.N.: *377*

Date: *SEPT 22 80*



50 VOLTS/cm
2.0 nsec/cm
1.0 MHz

- a) Output signal amplitude:
+67 VOLTS.
- b) Pulse width:
≤ 2 NSECS (10%)
- c) Jitter, stability:
OK
- d) Prime power:
+15 ± 1.0 VOLT
190 mA MAX

[Signature]