


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-N*

S.N.: *380*

Date: *SEPT 25 80*



- a) Output signal amplitude:
0 TO -10V
- b) Pulse width:
0.2 TO 4.0 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 130 psec
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT

10 VOLT/DIV
1.0 nsec/DIV
400 kHz.

250 mA MAX