


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI*

S.N.: *402*

Date: *NOV 14 80*



a) Output signal amplitude:

> +40 VOLTS

b) Pulse width:

0.6 TO 100 NSEC.

c) Rise time:

≤ 500 PSEC

d) Fall time:

≤ 2.0 NSEC.

e) PRF:

OK

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLT.

33 VOLTS/CM

1.0 NSEC/CM.

200 mA MAX.