


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AU-1*

S.N.: *405*

Date: *NOV 26 80*



*10.0 VOLTS/cm
1.0 nSEC/cm
400 kHz*

a) Output signal amplitude:

0 TO +10V

b) Pulse width:

0 TO 4.0 nSEC

c) Rise time:

≤ 100 pSEC

d) Fall time:

≤ 130 pSEC

e) PRF:

0 TO 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLT

300 mA MPN