


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-AV-P*  
S.N.: *409*  
Date: *Nov 28 80*



- a) Output signal amplitude:  
*0 TO  $\geq + 250 V$*
- b) Pulse width:  
*0 TO  $\geq 100 NSEC$*
- c) Rise time:  
 *$\leq 5 NSEC$*
- d) Fall time:  
 *$\leq 5 NSEC$*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24  $\pm$  1.0 VOLT*  
*450 mA MAX*

*10 NSEC / DIV*  
*330 VOLTS / DIV*  
*2.5 KHZ*