


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-MP*

S.N.: *411*

Date: *NOV 19 80*



*33 VOLTS / cm
1.0 NSEC / cm.
(400 KHz)*

a) Output signal amplitude:

> +40V

b) Pulse width:

0.5 TO 100 NSEC

c) Rise time:

≤ 500 psec

d) Fall time:

≤ 2.0 NSEC

e) PRF:

SEE FIG 2

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLT.

200 mA (max)