

IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT-2

S.N.: 414

Date: NOV 24 '80

AV<sup>0-3</sup>  
DS

a) Output signal amplitude:

$\geq +60 \text{ V}$

b) Pulse width:

$\leq 2.0 \text{ nSEC}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ V}$

190 mA MAX



33 VOLTS/cm

2.0 nSEC/cm