


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AU-1-N*

S.N.: *419*

Date: *DEC 12 1980*



- a) Output signal amplitude:  
*0 TO -10V*
- b) Pulse width:  
*0.2 TO 4.0 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 130 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1.0 VOLT*

*300 mA (MAX)*

*10 VOLTS/CM*  
*1.0 NSEC/CM*

*400 KHz*