


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-1*

S.N.: *432*

Date: *JAN 12 81*



a) Output signal amplitude:

b) Pulse width:

22 nsec

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 V

*200 mA
(mA)*

*64 Amps/cm
10 nsec/cm
1 kHz*