


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT-2

S.N.: 442

Date: JAN 28 81



a) Output signal amplitude:

$\geq +60 \text{ V}$

b) Pulse width:

$\leq 2.0 \text{ nSEC}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ V}$

150 mA (MAX)

32 VOLTS/cm

2.0 nSEC/cm

400 KHZ