


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-N*

S.N.: *448*

Date: *FEB 5 1981*



a) Output signal amplitude:

*> -15 VOLTS
(-23 VOLTS)*

b) Pulse width:

OK

c) Jitter, stability:

OK.

d) Prime power:

*+15 ± 1.0 VOLT
100 mA*

*0.5 nsec/cm.
10 VOLTS/cm.*