


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-P*
S.N.: *452*
Date: *FEB 13 1981*



- a) Output signal amplitude:
 $\geq +40V$
- b) Pulse width:
 $0.6 \text{ TO } 100 \text{ NSEC}$
- c) Rise time:
 $\leq 500 \text{ PSEC}$
- d) Fall time:
 $\leq 2 \text{ NSEC}$
- e) PRF:
 $0 \text{ TO } 1 \text{ MHz,}$
(SEE PL)
- f) Jitter, stability:
OK
- g) Prime power:

$+15 \pm 1.0 \text{ VOLT}$
 150 mA (MAX)

2 NSEC / CM
 32 VOLTS / CM
 40 KHz