


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-N*

S.N.: *453*

Date: *FEB 24 1981*



a) Output signal amplitude:

0 TO -15V

b) Pulse width:

0.2 TO 2.0 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 VOLT

600 mA (MAX)

*10 VOLT/cm
1.0 NSEC/cm
4.0 MHz*