


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT 2*

S.N.: *454*

Date: *FEB 18 81*



a) Output signal amplitude:

*≥ 60 VOLTS.*

b) Pulse width:

*≤ 2.0 NSEC.*

c) Jitter, stability:

*OK.*

d) Prime power:

*+15 ± 1.0 V*

*160 mA*

*32 VOLTS/DIV.*

*2.0 NSEC/DIV*

*400 KHZ*