


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT2

S.N.: 455

Date: FEB 18 81



a) Output signal amplitude:

$\geq 60$  VOLTS

b) Pulse width:

$\leq 2.0$  NSEC

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0$  V

165 mA

32 VOLTS/DIV

2.0 NSEC/DIV

400 KHZ