


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-A*

S.N.: *459*

Date: *FEB 24 '81*



*2 NSEC/DIV  
32 VOLTS/DIV  
40 KHZ*

- a) Output signal amplitude:  
*0 TO +40 V*
- b) Pulse width:  
*0.6 TO 100 NSEC*
- c) Rise time:  
*SEE FIG. 1A + FIG. 1*  
*≤ 0.5 NSEC*
- d) Fall time:  
*≤ 2.0 NSEC.*
- e) PRF:  
*SEE FIG 2.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 ± 1.0 VOLT*  
*150 mA (MAX)*