


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVK-10-N*

S.N.: *461*

Date: *FEB 27 81*



*10 NSEC/DIV
10 VOLTS/DIV*



*1.0 NSEC/DIV
10 VOLTS/DIV
20 KHZ*

- a) Output signal amplitude:
0 TO -15 V
- b) Pulse width:
0 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
*+15 ± 1.0 VOLT
300 mA (MAX)*

[Signature]