

PULSE GENERATOR
PERFORMANCE CHECK


Model: AVI-AV-P

S.N.: 466

Date: FEB 28 81



10 NSEC/DIV
32 VOLTS/DIV
20 KHZ



1.0 NSEC/DIV
32 VOLTS/DIV
20 KHZ

a) Output signal amplitude:

0 TO +50 V

b) Pulse width:

0 TO 100 NSEC

c) Rise time:

\leq 500 PSEC

d) Fall time:

\leq 2 NSEC

e) PRF:

0 TO 20 KHZ

f) Jitter, stability:

OK

g) Prime power:

+15 \pm 1.0 VOLT

200 mA (MAX)