


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-V-1-P*

S.N.: *472*

Date: *MAR 13 81*



- a) Output signal amplitude:
+ 10 V
- b) Pulse width:
0.2 TO 4.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 130 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK.
- g) Prime power:

*1.0 NSEC/DIV
10 VOLTS/DIV
400 KHZ*

*+15 ± 1.0 VOLT
325 mA (MAX)*