


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-V-1-P*

S.N.: *473*

Date: *MAR 13 81*



*1.0 nSEC/DIV*  
*10 VOLTS/DIV*  
*400 KHz*

- a) Output signal amplitude:  
*+ 10V*
- b) Pulse width:  
*0.2 TO 4.0 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 130 pSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1.0 VOLTS*  
*300 mA (MAX)*