


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-AV-P*

S.N.: *478*

Date: *MAR 23 81*



a) Output signal amplitude:

*0 TO +250V*

b) Pulse width:

*0 TO 100 NSEC*

c) Rise time:

*≤ 5 NSEC*

d) Fall time:

*≤ 5 NSEC*

e) PRF:

*0 TO 5 KHZ*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 10 VOLT*  
*400 mA (MAX)*

*320 VOLTS/CM*  
*10 NSEC/CM*  
*2.0 KHZ*