


PULSE GENERATOR
PERFORMANCE CHECK


Model: AVK-V-N-05

S.N.: 483

Date: APRIL 20 81



10 VOLTS/DIV
10 nSEC/DIV
20 KHZ



10 VOLTS/DIV
1.0 nSEC/DIV
20 KHZ

- a) Output signal amplitude:
-5 VOLTS
- b) Pulse width:
0 TO 50 nSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLTS
275 mA (MAX)