


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-1-A*

S.N.: *502*

Date: *MAY 5 81*



*100 V/DIV  
2.0 NSEC/DIV  
40 KHZ*

a) Output signal amplitude:

*0 TO +250 V*

b) Pulse width:

*≤ 4 NSEC*


c) Jitter, stability:

*OK*

d) Prime power:


*+15 ± 1.0 VOLTS*

*230 mA (MAX)*



*100 V/DIV  
2.0 NSEC/DIV  
40 KHZ*

*[Signature]*



*100 V/DIV  
0.5 NSEC/DIV  
40 KHZ*