


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-N*

S.N.: *505*

Date: *APRIL 29 81*



- a) Output signal amplitude:
-5 VOLTS
- b) Pulse width:
4 TO 100 NSECS
- c) Rise time: *SEE FILE 142*
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF: *OK SEE FILE 142*
- f) Jitter, stability:
OK
- g) Prime power:

2.0 NSEC/DIV
10 VOLTS/DIV
40 KHZ

+15 ±1.0 VOLT
210 MA MAX