


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2A-W-PS

S.N.: 506

Date: June 4 81



100 Volts/DIV  
10 nsec/DIV  
4 kHz.

- a) Output signal amplitude:  
0 TO +180 V
- b) Pulse width:  
0 TO 400 nsec
- c) Rise time:  
≤ 1 nsec
- d) Fall time:  
52 nsec FWH
- e) PRF: PW < 100 nsec  
0 TO 5 kHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120 V 60 Hz

*[Handwritten signature]*