


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-AV-N-PS*

S.N.: *507*

Date: *MAY 25 81*



*320 VOLTS/DIV  
10 NSEC/DIV  
2 KHZ.*

- a) Output signal amplitude:  
*0 TO -250 V*
- b) Pulse width:  
*0 TO 100 NSEC*
- c) Rise time:  
*≤ 5 NSEC*
- d) Fall time:  
*≤ 5 NSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+120 VAC  
60 HZ.*

*[Signature]*