


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AV14-AM 2

S.N.: 510

Date: APRIL 28 81



a) Output signal amplitude:

+ 60 V

b) Pulse width:

≤ 2.0 nsec

c) Jitter, stability:

OK

d) Prime power:

+135 \pm 5.0 volts

4 mA (MAX)

50 VOLTS/DIV
2.0 nsec/DIV
1.0 MHz