


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-1VA*

S.N.: *512*

Date: *APRIL 28 81*



*100 VOLTS/DIV*

*2.0 NSEC/DIV*

*40 KHZ*

a) Output signal amplitude:

*0 to +90V*

b) Pulse width:

*≤ 1.0 nsec*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 0.5V*

*190 mA*

e) PRF:

*0 TO 40 KHZ*