


IMPULSE GENERATOR

PERFORMANCE CHECK


Model: AVH-5-1-N

S.N.: 518

Date: 11/29/81



0.1 NSC/1DIV
10 V OBT/1DIV



1.0 NSC/1DIV
10 V OBT/1DIV
400 KHZ.

a) Output signal amplitude:

-10V

b) Pulse width:

≤ 150 PSCT

c) Jitter, stability:

OK.

d) Prime power:

+15 ± 1.0 V

180 mA MAX

442-