


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVP-AV-1**

S.N.: **519**

Date: **MAY 29 81**



1.0 nsec/div  
10 volts/div  
400 kHz

- a) Output signal amplitude:  
**0 TO +10V**
- b) Pulse width:  
**0 TO 4 nsec**
- c) Rise time:  
**≤ 100 psec**
- d) Fall time:  
**≤ 135 psec**
- e) PRF:  
**0 TO 1.0 MHz**
- f) Jitter, stability:  
**OK**
- g) Prime power:  
**+15 ± 1.0 V**  
**275 mA (MAX)**

*[Signature]*