


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVK-AV*

S.N.: *525*

Date: *MAY 25 81*



*10 NSEC/DIV
10 VOLTS/DIV
20 KHZ*



*1.0 NSEC/DIV
10 VOLTS/DIV
20 KHZ* } TR.

a) Output signal amplitude:

0 TO +15V

b) Pulse width:

0 TO 50 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 50 KHZ

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

250 mA (MAX)

[Signature]